



المعهد الوطني للمعايير والتقنية
Emirates Authority for Standardization and Metrology



Certificate of Accreditation

EMIRATES NATIONAL ACCREDITATION SYSTEM

(has accredited)

METROMAC ACCREDITATION CALIBRATION LAB

its Calibration Laboratory

NAL/004

Located in Masfout, Abu Dhabi, UAE

The Laboratory complies with the requirements of the Standard ISO/IEC 17025:2005

'General Requirements for the Competence of Testing and Calibration Laboratories' in the Emirate (E.A.A) Conditions for the calibration specified in the attached accreditation scope. The accreditation covers regular surveillance.

The organization itself is responsible for the consistent performance evaluation.

Date of Initial Accreditation was 08/04/2016

Date of Renewal: 08/04/2018

Valid Until: 07/04/2020

Emirates National Accreditation System


Dr. Rashid Ahmad Bin Fahed
Minister of Environment and Water
Chairman of EEMA



02 651 2000 (Main) 02 651 2001 (Fax)

02 651 2002 (Toll Free) 02 651 2003 (Emergency)

02 651 2004 (Toll Free) 02 651 2005 (Toll Free)

02 651 2006 (Toll Free) 02 651 2007 (Toll Free)



02 651 2008 (Main) 02 651 2009 (Fax)

02 651 2010 (Toll Free) 02 651 2011 (Toll Free)

02 651 2012 (Main) 02 651 2013 (Fax)

02 651 2014 (Toll Free) 02 651 2015 (Toll Free)

Accreditation Scope

METRO MAC AUTOMATION Calibration Laboratory, NAL 004
Masrafah, Abu Dhabi, UAE

#	Measurand	Measuring Range	CMC (±)	Description of Method and Equipment Used
Electrical				
1	DC Voltage	10 μ V to 100 mV	0.002 %	Direct comparison of generated DC voltage from calibrator using METRO MAC standard Calibration procedure. Equipment Used : Multi-product Calibrator Fluke 5220A Serial No: 9575083
		100 mV to 1 V	0.002 %	
		1 V to 10 V	0.002 %	
		10 V to 100 V	0.002 %	
		100 V to 1000 V	0.002 %	
2	AC Voltage	1 mV to 10 mV	0.04 %	Direct comparison of generated AC voltage from calibrator using METRO MAC standard Calibration procedure. Equipment Used : Multi-product Calibrator Fluke 5220A Serial No: 9575083
		50 Hz to 10 kHz		
		100 Hz to 100 kHz	0.04 %	
		10 mV to 100 mV	0.03 %	
		50 Hz to 10 kHz		
		10 kHz to 100 kHz	0.1 %	
		100 mV to 1 V	0.03 %	
		50 Hz to 10 kHz		
		10 kHz to 100 kHz	0.05 %	
		1 V to 10 V	0.03 %	
		50 Hz to 10 kHz		
		10 kHz to 100 kHz	0.05 %	
10 V to 100 V	0.03 %			
50 Hz to 10 kHz				
10 kHz to 100 kHz	0.05 %			
100 V to 1000 V	0.04 %			
50 Hz to 100 kHz				
3	DC Current	20 μ A to 1 mA	0.02 %	Direct comparison of generated DC current
		1 mA to 100 mA	0.02 %	
		100 mA to 1 A	0.02 %	
		1 A to 10 A	0.05 %	From calibrator using METRO MAC standard Calibration procedure. Equipment Used : Multi-product Calibrator Fluke 5220A Serial No: 9575083
		10 A to 20 A	0.1 %	
		10 A to 1000 A	0.1 %	Comparison to Calibrator 9520A passing current through 50 amp current coil

Accreditation Scope

METRO MAC AUTOMATION Calibration Laboratory, NAL 004
Musafah, Abu Dhabi, UAE

#	Measured	Measuring Range	CMC (±)	Description of Method and Equipment Used
4	Resistance	1 Ω to 180 Ω	0.1%	Direct comparison of shunted resistance from calibrator using METRO MAC standard Calibration procedure. Equipment used: Multi-product Calibrator Fluke 5820A Serial No: 9875003
		0.1 kΩ to 18 kΩ	0.008%	
		18 kΩ to 180 kΩ	0.008%	
		0.1 MΩ to 1.8MΩ	0.005%	
		1 MΩ to 18 MΩ	0.005%	
10 MΩ to 300 MΩ	0.005%			
5	AC Current	50 μA to 100 μA 50 Hz to 1 kHz	0.2%	Direct comparison of generated AC current from calibrator using METRO MAC standard Calibration procedure. Equipment used: Multi-product Calibrator Fluke 5820A Serial No: 9875003
		0.1 mA to 1 mA 50 Hz to 10 kHz	0.2%	
		1 mA to 10 mA 50 Hz to 10 kHz	0.005%	
		10 mA to 2 A 50 Hz to 1 kHz	0.01%	
		2 A to 10 A 50 Hz to 1 kHz	0.1%	
		10 A to 20 A 50 Hz to 1 kHz	0.1%	
		10 A to 1800 A Frequency 45-65 Hz	0.15%	
Temperature				
6	Digital Thermometers (RTD /Thermocouple Sensors with Indicator) (Repeatability 0.01°C)	-40 °C to 100°C	0.5°C	-Comparison Method by using: 1- Temperature Calibrator, Armetek SI No : 541828-08743 2- Calibration Bath, SI No:5034-85 3- SPRT Probe Pt 100,SI No : 808771 4- Digital Multimeter Fluke SI No : 9422818
	Digital Thermometers (RTD /Thermocouple Sensors with Indicator) (Repeatability 0.01°C)	180 °C to 420°C	1.1°C	

Accreditation Scope

METRO MAC AUTOMATION Calibration Laboratory, MAL 004
Mussafah, Abu Dhabi, UAE

#	Measured	Measuring Range	CMC (k=1)	Description of Method and Equipment Used
7	Oven		1°C	Comparison Method by using 1- Digital Temperature Data logger (Kotagawa) 2- Flexible PRT Probes pt 180 3- Multifunction-Calibrator Check, TR3
	Freezer/ Chiller		0.01°C	
	Insulator	-20°C to 250°C	0.01°C	
	Water Bath		0.01°C	
Pressure				
8	Hydraulic Pressure Gauges Low Range	1 to 60 bar	0.00 bar	Direct comparison with pressure balanced by standard weights applying force on a piston cylinder assembly. Equipment used Ozer Weight Tester DH 8U02H0103 180H0A, Piston Cylinder GH 374H
9	Hydraulic Pressure Gauges High Range	80 to 700 bar 700 to 1000 bar	0.10 bar 0.4 bar	Direct comparison with pressure balanced by standard weights applying force on a piston cylinder assembly. Equipment used Ozer Weight Tester DH 8U02H0103 180H0A, Piston Cylinder GH 374H
10	Pneumatic Pressure Gauges	0.005 bar to 20 bar	0.02 bar	Direct comparison with pneumatic pressure calibrator Equipment used Pressure-Calibrator OZEUX DP1801 GH 9450103-3 & PAGE 1085
11	Vacuum Gauges	-0.008 to -0.8 bar	0.021 bar	Direct comparison using vacuum source and a standard pressure indicator. Equipment used Precision Pressure Indicator OZEUX DP1 105 GH 10002198



Accreditation Scope

METRO MAC AUTOMATION Calibration Laboratory, MAL 004
Mussafah, Abu Dhabi, UAE

#	Measurement	Measuring Range	CMC (±)	Description of Method and Equipment Used
Dimensional Length				
12	Vernier Calipers Repeatability 0.01 mm	1mm to 150 mm 1mm to 200 mm 1mm to 300 mm 1mm to 600 mm	10 µm 10 µm 10 µm 10 µm	Comparison Method 1- Set of gauge Blocks Tesa, SI No: 2018-320 2- 125 mm Block Tesa, SI No: 86117 3- 200 mm Block Tesa, SI No: 87260 4- 300 mm Block Tesa, SI No: 87143 5- Caliper Checker Actya, SI No: 823 6- 25mm, 50mm, 75mm, 100mm- Gauge Blocks Mako, Tesa, SI No: 15-30080, 1-30070, 15-30058, 2-30085
13	Outside Micrometers, Repeatability 0.001 mm	1 to 25 mm 25 to 50 mm 50 to 75 mm 75 to 100 mm	2 µm 2 µm 2 µm 2 µm	Comparison Method 1- Set of gauge Blocks Tesa SI no: 2018-320 2- 125 mm Block Tesa SI No: 86117 3- 200 mm Block Tesa SI No: 87260 4- 300 mm Block Tesa SI No: 87143
14	Outside Micrometers Repeatability 0.01 mm	1 mm to 150 mm 150mm to 300 mm	10 µm 10 µm	5- Set of Slip Gauges Actya, SI No: 2695 6- 25mm, 50mm, 75mm, 100mm- Gauge Blocks Mako, Tesa, SI No: 15-30080, 1-30070, 15-30058, 2-30085
Volume				
15	Micropipettes	1µl - 10 µl	0.00µl	By using Automatic Pipette Station having Ambient Conditions Monitor Measuring Module (Microbalance)
	Micropipettes	10µl - 100 µl	0.00µl	
	Micropipettes	100µl - 1000 µl	0.50µl	
	Micropipettes	1000µl - 12000 µl	0.70µl	



Accreditation Scope

METRO MAC AUTOMATION Calibration Laboratory, NAL 004
Mussafah, Abu Dhabi, UAE

#	Measured	Measuring Range	CMC (k=1)	Description of Method and Equipment Used
Mass				
10	Digital Weighing Balance	10 - 21 g Readability 0.001 mg	0.08 mg	Comparison Method By using E2 Class Weights Set
	Digital Weighing Balance	5 - 220 g Readability 0.01 mg	0.24 mg	
	Digital Weighing Balance	5 - 1000 g Readability 0.001g	0.0018 g	
11	Digital Weighing Balance	0 - 5200 g Readability 0.01g	0.03 g	Comparison Method By using E2 Class Weights Set & F1 Class Weight
	Digital Weighing Balance	0 - 20 kg Readability 0.5g	0.64 g	
18	Digital Weighing Balance	0 - 180 kg Readability 10 g	16.44 g	Comparison Method By using E2 Class Weights Set, F1 Class Weights & M1 Class Weights
	Digital Weighing Balance	0 - 600 kg Readability 50 g	20.75 g	
END				
				
 Program's Manager Signature				